


Allan Deviation $\sigma_y(\tau)$



Allan Deviation $\sigma_y(\tau)$

$\tau_0 = 100 \text{ ms}$	Avg. Time (s)	Allan Deviation $\sigma_y(\tau)$	NEQ BW = 5
	0.1	3.945×10^{-12}	
	0.2	4.679×10^{-12}	
	0.4	5.949×10^{-12}	
	1	4.286×10^{-12}	
	2	3.794×10^{-12}	
	4	3.649×10^{-12}	
	10	3.729×10^{-12}	
	20	3.85×10^{-12}	
	40	3.95×10^{-12}	
	100	4.38×10^{-12}	
	200	4.91×10^{-12}	
	400	5.30×10^{-12}	
	1000	6.2×10^{-12}	
	2000	6.2×10^{-12}	
	4000	6.8×10^{-12}	
	10000	7.7×10^{-12}	
	20000	1.5×10^{-11}	


27/05/2007 17:21:00
16h 19m

TSC 5120A

Phase Difference

2.0x10⁻⁰⁹ s/div

Center: -2.0734x10⁻⁰⁷ s



60s/div

Input 5.0 MHz 15 dBm

Reference 5.0 MHz 13 dBm


27/05/2007 17:21:00
16h 19m

TSC 5120A

Frequency Difference

8.0×10^{-12} /div

Center: 2.315×10^{-11}



60s/div

Input 5.0 MHz 15 dBm

Reference 5.0 MHz 13 dBm

Frequency Counter

Sample Time (s)	Frequency (MHz)
1	5.0000000016250
10	5.00000000168224
100	5.000000001716547
1000	5.000000001661944

Reference Frequency: 5.0 MHz (auto)

$\mathcal{L}(f)$ Phase Noise at 5.0 MHz (dBc/Hz)

20 dB/Div

Mx: 0.976562 Hz Δ Mx: 10009.77 Hz Δ x: -10008.8
 My: -99.4 dBc/Hz Δ My: -136.1 dBc/Hz Δ y: 36.7

