


Allan Deviation $\sigma_y(\tau)$



Allan Deviation $\sigma_y(\tau)$

$\tau_0 = 1 \text{ ms}$	Avg. Time (s)	Allan Deviation $\sigma_y(\tau)$	NEQ BW = 500
	0.001	3.6145×10^{-11}	
	0.002	2.5427×10^{-11}	
	0.004	1.7852×10^{-11}	
	0.01	1.1057×10^{-11}	
	0.02	7.862×10^{-12}	
	0.04	5.592×10^{-12}	
	0.1	3.552×10^{-12}	
	0.2	2.555×10^{-12}	
	0.4	2.027×10^{-12}	
	1	1.985×10^{-12}	
	2	3.142×10^{-12}	
	4	5.29×10^{-12}	
	10	9.62×10^{-12}	
	20	1.387×10^{-11}	
	40	1.89×10^{-11}	
	100	3.41×10^{-11}	
	200	6.2×10^{-11}	
	400	1.20×10^{-10}	
	1000	3.0×10^{-10}	
	2000	6.0×10^{-10}	
	4000	1.21×10^{-9}	
	10000	3.1×10^{-9}	


16/06/2007 19:37:20
6h 34m

TSC 5120A

Phase Difference

7.0x10⁻⁰⁷ s/div

Center: -1.18802x10⁻⁰⁴ s



60s/div


Input 5.0 MHz 14 dBm

Reference 5.0 MHz 15 dBm

Frequency Difference

4.0×10^{-11} /div

Center: -9.7563×10^{-09}



60s/div

Input 5.0 MHz 14 dBm

Reference 5.0 MHz 15 dBm

Frequency Counter


Sample Time (s)	Frequency (MHz)
1	4.9999999366370
10	4.99999993666017
100	4.999999936811534
1000	4.999999935910069

Reference Frequency: 5.0 MHz (auto)

$\mathcal{L}(f)$ Phase Noise at 5.0 MHz (dBc/Hz)

Mx: 0.976562 Hz Δ Mx: 99975.59 Hz Δ x: -99974.6
 My: -114.5 dBc/Hz Δ My: -170.2 dBc/Hz Δ y: 55.8

25 dB/Div



Input 5.0 MHz 14 dBm

Reference 5.0 MHz 15 dBm